LARGE RANGE DIAL TEST INDICATOR

JEWELED BEARING

INSIZE MANUFACTURER INSPECTION CERTIFICATE TRACEABLE TO NIST

2386-16A



clamps



- Dial Reading: 0-40-0
- Jeweled bearing
- Carbide contact point
- Two measuring directions
- Anti magnetic body
- Supplied with two clamps: diameter Ø4mm and Ø8mm
- Optional accessory: styli (page 106)

Code	Range	Graduation	Accuracy	Hysteresis
2386-16A*	1.6mm	0.01mm	25µm	8µm

*Supplied with manufacturer inspection certificate traceable to NIST USA