



- With limit pointers
- Optional accessory: backs (page 101), contact points (page 98~100)



2889-30

Code	Range	Graduation	Accuracy	Hysteresis	Range/rev	Dial reading	Remark
2889-30*	30mm	0.01mm	25 μ m	5 μ m	1mm	0-100	lug back
2889-30F*	30mm	0.01mm	25 μ m	5 μ m	1mm	0-100	flat back

*Supplied with manufacturer inspection certificate traceable to NIST USA

DIAL INDICATOR

Unit: mm

