



2892-10

- With limit pointers
- Optional accessory: backs (page 101), contact points (page 98~100)

Code	Range	Graduation	Accuracy	Hysteresis	Range/rev	Dial reading	Remark
2892-10*	10mm	0.01mm	17 μ m	3 μ m	1mm	0-100	lug back
2892-10F*	10mm	0.01mm	17 μ m	3 μ m	1mm	0-100	flat back

*Supplied with manufacturer inspection certificate traceable to NIST USA

DIAL INDICATOR

