BACK PLUNGER LONG STYLI DIAL TEST INDICATOR







. (Cera	amic	bearing	1

- Anti magnetic carbide contact point
- Two measuring directions
- Supplied with one clamp: diameter Ø8mm
- Optional accessory: styli (page 106)

Code	Range	Graduation	Accuracy	Hysteresis	Dial reading
2899-05*	0.5mm	0.01mm	13um	3um	0-25-0

*Supplied with manufacturer inspection certificate traceable to NIST USA

